



Evaluation Board Documentation

ADE7754 Energy metering IC

Preliminary Technical Data

EVAL-ADE7754EB

FEATURES

Evaluation Board is designed to be used together with accompanying software to implement a fully functional Three-Phase Energy Meter (Watt-Hour Meter).

Easy connection of various external transducers via screw terminals.

Easy modification of signal conditioning components using PCB sockets.

LED indicators on logic outputs CF, VARCF, and IRQ.

Optically isolated data output connection to PC parallel port.

Optically isolated frequency output (CF) to BNC.

External Reference option available for on-chip reference evaluation.

GENERAL DESCRIPTION

The ADE7754 is high accuracy electrical active power measurement ICs for three-phase applications with a pulse output. This output is intended to be used for calibration purposes. The ADE7754 incorporate ADCs, reference circuitry, and all the signal processing required to perform active power and energy measurement.


The ADE7754 provides Active Energy, RMS values, temperature measurement and Apparent Energy information via a serial interface. This documentation describes the ADE7754 evaluation kit Hardware and Software functionality. The evaluation board (watt-hour meter) is configured via the parallel port of a PC. The data interface between the

evaluation board and the PC is fully isolated. Windows™ based software is provided with the evaluation board which allows it to be quickly configured as an energy meter.

The ADE7754 evaluation board and this documentation, together with the ADE7754 data sheet provide a complete evaluation platform for the ADE7754.

The evaluation board has been designed so that the ADE7754 can be evaluated in the end application, i.e., Watt-Hour Meter. Using the appropriate transducers on the current channel (e.g., CT) the evaluation board can be connected to a test bench or high voltage (240V rms) test circuit. On-board resistor dividers networks provide the attenuation for the line voltages. This application note also describes how the current transducers should be connected for the best performance. The evaluation board requires two external 5V power supplies (one is required for isolation purposes) and the appropriate current transducers.

FUNCTIONAL BLOCK DIAGRAM



REV. PrD 01/03

Information furnished by Analog Devices is believed to be accurate and reliable. However, no responsibility is assumed by Analog Devices for its use, nor for any infringements of patents or other rights of third parties which may result from its use. No license is granted by implication or otherwise under any patent or patent rights of Analog Devices.

One Technology Way, P.O. Box 9106, Norwood, MA 02062-9106, U.S.A.
Tel: 617/329-4700 Fax: 617/326-8703

EVAL-ADE7754EB

ANALOG INPUTS (P4, P5, P6, P7, P8 AND P11)

Voltage and current signals are connected at the screw terminals P4-P6 and P7-P8 and P11 respectively. All analog input signals are filtered using the on-board anti-alias filters before being presented to the analog inputs of the ADE7754. The default component values which are shipped with the evaluation board are the recommended values to be used with the ADE7754. The user can easily change these components, however this is not recommended unless the user is familiar with sigma-delta converters and also the criteria used for selecting the component values for the analog input filters - See ADE7754 datasheet.

Current sense inputs (P4, P5 and P6)

P4, P5 and P6 are two-way connection blocks which allow ADE7754's current inputs of phase A, B and C respectively to be connected to current transducers. Figure 1 shows the connector P4 and the filtering network which is provided on the evaluation board.

The resistors SH1A, SH2A, SH1B, SH2B, SH1C and SH2C are by default not populated. They are intended to be used as burden resistors when CTs are used as the current transducers—see using a CT as a the current transducer.

The RC networks R9/C5, R7/C8, R15/C9, R13/C12, R21/C13, R19/C16 are used to provide phase compensation when a Current Transformer is being used as the current transducer with the ADE7754—see using a Current Transformer as the current transducer. These RC networks are easily disabled by placing JP4, JP1, JP10, JP7, JP16 & JP13 and removing C5, C8, C9, C12, C13 and C16 (socketed).

The RC networks R10/C6, R8/C7, R16/C10, R14/C11, R22/C14 and R20/C15 are the anti-alias filters which are required by the on-chip ADCs. The default corner frequency for these LPFs (Low Pass Filters) is selected as 4.8kHz (1k Ω & 33nF). These filters can easily be adjusted by replacing the components on the evaluation board. However before adjusting the component values the user should first review the ADE7754 datasheets.




Figure 1 — Current Channel on the ADE7754 evaluation board

Using a CT as the current transducer

Figure 2 shows how a CT can be used as a current transducer in one phase of a 3-phase 4-wire distribution system (Phase A). In a three phase distribution system Phase A, Phase B and Phase C are nominally 120° phase difference to each other. Each phase usually requires a connection of this type for current sensing.




Figure 2 — CT connection to Current Channel

The CT secondary current is converted to a voltage by using a burden resistance across the secondary winding outputs. Care should be taken when using a CT as the current transducer. If the secondary is left open, i.e., no burden is connected, a large voltage could be present at the secondary outputs. This can cause an electrical shock hazard and potentially damage electronic components.



Warning!

Using a CT without a burden resistor can lead to electrical shock.

The anti-alias filters should be enabled by opening jumpers JP5, JP6, JP11, JP12, JP17 and JP18—see Figure 2.

Most CTs will have an associated phase shift of between 0.1° and 1° at 50Hz/60Hz. This phase shift or phase error can lead to significant energy measurement errors, especially at low power factors. However this phase error can be corrected by writing to the Phase Calibration registers (APHCAL, BPHCAL and CPHCAL) in the ADE7754. The software supplied with the ADE7754 evaluation board allows user adjustment of the Phase Calibration register. See the *Evaluation Software Description* for more information.

For this example, notice that the maximum analog input range on the Current channel is set to 0.5V.

When using the ADE7754, the Gain for Current channel has been set to 1. The maximum analog input range and gain are set via the Gain register (GAIN)—see the ADE7754 data sheet. The evaluation software allows the user to configure the current channel gain. This means that the maximum peak differential signal on the current channel is 0.5V.

PRELIMINARY TECHNICAL DATA

EVAL-ADE7754EB

Voltage sense inputs (P7, P8 and P11)

The voltage inputs connections on the ADE7754 evaluation board can be directly connected to the line voltage sources. The line voltages are attenuated using a simple resistor divider network before it is presented to the ADE7754. The attenuation network on the voltage channels is designed such that the corner frequency (3dB frequency) of the network matches that of the RC (anti-aliasing) filters on the current channels inputs. This is important, because if they do not match there will be large errors at low power factors.

The Serial interface along with the digital parameters of the ADE7754 enable a fully software calibration procedure. In this case, the voltage attenuation networks are made up of R48, R49 and R47 for Phase A, R45, R46 and R44 for Phase B and R41, R42 and R43 for Phase C. Figure 3 shows a typical connection for the line voltage when using fixed resistors.




Figure 3 — Phase A Voltage Channel attenuation network using Fixed resistors

The attenuation networks can be easily modified by the user to accommodate any input signal level. However the value of R43, R44 and R47 (1kΩ) should not be altered as the phase response of Voltage channels should match that of the current channels.

The maximum signal level permissible at VAP, VBP and VCP is 0.5V peak for the ADE7754. Although the ADE7754 analog inputs can withstand $\pm 6V$ without risk of permanent damage, the signal range should not exceed $\pm 0.5V$ with respect to AGND for the ADE7754 for specified operation. Note that the analog input VN is connected to AGND via the anti-alias filter R40/C1 using JP28. Jumper JP27 should be left open.

EVAL-ADE7754EB**TABLE II : JUMPER SETTINGS**

JUMPER	OPTION	DESCRIPTION
JP1-JP7-JP13	Closed	This will short out R7, R13 and R19 respectively. The effect is to disable the phase compensation filter on the analog inputs IAN, IBN and ICN respectively. Default Closed.
JP2-JP8-JP14	Closed	This will connect the analog input IAP, IBP and ICP respectively to ground. Default Open.
JP3-JP9-JP15	Closed	This will connect the analog input IAN, IBN and ICN respectively to ground. Default Open.
JP4-JP10-JP16	Closed	This will short out R9, R15 and R21 respectively. The effect is to disable the phase compensation filter (for Current transformers) on the analog inputs IAN, IBN and ICN respectively. Default Closed.
JP5-JP11-JP17	Closed	This will short out R10, R16 and R22 respectively. The effect is to disable the anti-alias filter on the analog inputs IAP, IBP and ICP respectively. Default Open.
	Open	Enable the anti-alias filter on IAP, IBP and ICP respectively.
JP6-JP12-JP18	Closed	This will short out R8, R14 and R20 respectively. The effect is to disable the anti-alias filter on the analog inputs IAN, IBN and ICN respectively. Default Open.
	Open	Enable the anti-alias filter on IAN, IBN and ICN respectively.
JP19	Closed	This will connect the Analog and Digital ground planes of the PCB. Default Closed.
JP20	Closed	This will connect an external reference 2.5V (AD780) to the ADE7752/54
	Open	This will enable the ADE7754 on-chip reference.
JP21	Closed	This connects the DVDD and +5V (buffers) supply for the evaluation board together. Default Closed.
JP22	Closed	This connects the AVDD and DVDD supply for the evaluation board together. Default Closed.
JP23-JP24-JP25	Closed	This will short the attenuation network on the voltage channels. Default Open
JP26	Closed	This will connect the optical isolator ground to the evaluation board ground (DGND). If full isolation between the evaluation board and PC is required, this jumper should be left open
JP27	Closed	This will short out R40. The effect is to disable the anti-alias filter on the analog input VN. Default Open
JP28	Closed	This will connect the analog input VN to ground. Default Closed.
JP29-JP35-JP38	Closed	This will connect the analog inputs VCP, VBP and VAP to ground respectively. Default Open-
JP47-JP48-JP49	Closed	This will short out disconnect Analog input VCP, VBP and VAP respectively from the ADE7754. Default Closed
JP32	Closed	This will short out disconnect Analog input VN from the ADE7754. Default Closed
JP43	A	This connects the buffered logic output CF to BNC J3 connector via an optical isolator.
	B	This connects the buffered logic output CF to the LED CR2.
JP44	A	This connects the buffered logic output VARCF to BNC J1 connector via an optical isolator.
	B	This connects the buffered logic output VARCF to the LED CR1.

PRELIMINARY TECHNICAL DATA

EVAL-ADE7754EB

SETTING UP THE EVALUATION BOARD AS AN ENERGY METER

Figure 4 shows a typical set up for the ADE7754 evaluation board. In this example a kWh meter for a 4 wire, three phase distribution system is shown. Current Transformers are used to sense phases' current and are connected as shown in Figure 4. For a more detailed description on how to use a CT as a current transducer see the *Current Sense Inputs* section of this documentation. The line voltage is connected directly to the evaluation board as shown. Note JP23, JP24 and JP25 should be left open to ensure that the attenuation networks are not bypassed. Also note the use of two power supplies.

The second power supply is used to power the optical isolation. With JP26 left open, this will ensure that there is no electrical connection between the high voltage test circuit and the frequencies outputs. The power supplies should have floating voltage outputs.

Setting up the evaluation board for the ADE7754

The configuration of the jumpers on the ADE7754 evaluation board for operation with the ADE7754 is detailed in Figure 4.

The evaluation board is connected to the PC parallel port using the cable supplied. The cable length should not exceed 6 feet (2 meters) or the serial communication between the PC and the evaluation board may become unpredictable and error prone.

JP46 is provided to prevent the reset of the ADE7754 when the parallel cable is disconnected. When the parallel cable is plugged to the ADE7754 evaluation board, JP46 can be left open. Then the ADE7754 can be programmed via the parallel cable and the evaluation software.

If the parallel cable needs to be disconnected and the ADE7754 parameters retained, JP46 should be close before operation and stays in this position until the parallel cable is reconnected.

When the evaluation board has been powered up and is connected to the PC, the supplied software can be launched. The software will automatically start in default mode. The next section describes the ADE7754 evaluation software in detail and how it can be installed and uninstalled.




Figure 4 - Typical set up for the ADE7752-ADE7754 evaluation board

EVAL-ADE7754EB

ADE7754 EVALUATION SOFTWARE

The ADE7754 evaluation board is supported by Windows based software which will allow the user to access all the functionality of the ADE7754. The software is designed to communicate with the ADE7754 evaluation board via the parallel port of the PC.

Installing the ADE7754 Software

The ADE7754 Software is supplied on one CD-ROM. The minimum requirements for the PC are Pentium II 233MHz, 32 MB RAM, 10MB free HD space and at least one PS/2 or ECP parallel port. To install the software place the CD-ROM in the CDROM reader and double click "setup.exe". This will launch the set up program which will automatically install all the software components including the uninstall program and create the required directories. When the set up program has finished installing the "ADE7754Eval" program the user will be prompted to install the National Instruments run-time engine. This software was developed using National Instruments LabView software and the run-time engine is required in order to run the "ADE7754Eval" program. Follow the on-screen instructions to complete the installation. You will need to reboot your computer to complete the installation.

To launch the software simply go to the Start—>Programs—>ADE7754 menu and click on "ADE7754Eval".

Uninstalling the ADE7754 Evaluation Software

Both the "ADE7754Eval" program and the NI run-time engine are easily uninstalled by using the Add/Remove Programs facility in the control panel. Simply select the program to uninstall and click the Add/Remove button.

When installing a new version of the ADE7754 evaluation software, the previous version should be uninstalled.




Figure 5 — Uninstalling the ADE7754 Eval Software

Default Mode

When the software is launched, a message at the bottom of the window informs the user if the communication is possible between the PC and the evaluation board and which version of the ADE7754 is installed on the evaluation board. The user has the choice in different modes of operation: Active Energy, Apparent Energy Measurements, Voltage channels, Current Channels, Temperature and Interrupt. Each mode gives a view of the functionality of the ADE7754 and the Calibration mode gives the opportunity to use the eval. software like a true 3-phase Active Energy meter. This software should be used in conjunction with the ADE7754 datasheet.




Figure 6 - Default Window

Active Energy Measurement

In this mode, all the parameters needed for the measurement of the Active Energy are presented in one window - see Figure 7.




Figure 7 - Active Energy Measurement Window

The different modes, configuration and register values for the Active Energy Measurement can be selected in this window. All configuration registers are written or read to the ADE7754 when pushing the Write Configuration or Read Configuration buttons. The Active Energy and Line Active Energy registers are available by pushing the appropriate buttons. The Reactive Energy can also be accumulated over LINCYC half line cycles by setting the active or reactive switch and reading the Line Active Energy register.

PRELIMINARY TECHNICAL DATA

EVAL-ADE7754EB

The ADE7754 can be reset to its default configuration by pushing the reset button.

Apparent Energy Measurement

In this mode, all the parameters needed for the measurement of the Apparent Energy are presented in one window - see Figure 8.




Figure 8 - Apparent Energy Measurement Window

The different modes, configuration and registers values for the Apparent Energy Measurement can be selected in this window. All configuration registers are written or read to the ADE7754 when pushing the Write Configuration or Read Configuration buttons. The Apparent Energy and Line Apparent Energy registers are available by pushing the appropriate buttons.

The ADE7754 can be reset to its default configuration by pushing the reset button.

Voltage Channels

In this mode, all the parameters needed for the measurement of the Voltage channels are presented in one window - see Figure 9.




Figure 9 - Voltage Channels Window

The register values and information available for the Voltage Channels can be viewed in this window. All configuration registers are written or read to the ADE7754 when pushing the Write Configuration or Read Configuration buttons. To get the Voltage RMS, the "Read RMS value" should be depressed. Pressing "Read Period" displays the line Period for the selected phase. SAG, Peak and Zero-crossing Time-out interrupts information is available by pushing the "Read Events with Reset" button.

The Voltage Channel window gives also access to the Waveform sampling data. The sampling rate in this mode is 3.3ksps (CLKIN/3072). The user can define the number of samples and select the Voltage Channel Waveform to transfer.

The ADE7754 can be reset to its default configuration by pushing the reset button.

Current Channels

In this mode, all the parameters needed for the measurement of the Current channels are presented in one window - see Figure 10.




Figure 10 - Current Channels Window

The register values and information available for the Current Channels can be viewed in this window. All configuration registers are written or read to the ADE7754 when pushing the Write Configuration or Read Configuration buttons. The Current RMS and Peak interrupts information are available by pushing the appropriate buttons.

The Current Channel window gives also access to the Waveform sampling data. The sampling rate in this mode is 3.3ksps (CLKIN/3072). The user can define the number of samples and select the Current Channel Waveform to transfer.

The ADE7754 can be reset to its default configuration by pushing the reset button.

EVAL-ADE7754EB

Temperature

This window gives access to the temperature register of the ADE7754 - see Figure 11.




Figure 11 - Temperature Window

Interrupts

This window gives access to the Interrupt Mask and STATUS registers of the ADE7754 - see Figure 12.



Figure 12 - Interrupts Window

On line help

The ADE7754 evaluation software also comes with on-line help features. In order to activate the help function go to Help on the Menu Bar and select "Show Help"—see Figure 13. A Help Window will open. In order to get a description of a particular option (e.g., button, text box etc.) move the cursor over the item of interest. The Help window will display a description of the selected item.




Figure 13—On-line Help Function.

PRELIMINARY TECHNICAL DATA

EVAL-ADE7754EB

Measuring CT Phase Errors using the ADE7754

The ADE7754 itself can be used to measure a CT (and external components) phase error during calibration. The assumption is that the ADE7754 has no internal phase error (APHCAL = BPHCAL = CPHCAL = 00 hex) and the error due to external components is small (<0.5°). The procedure is based on a two point measurement, at PF=1 and PF = 0.5 (lag). The PF is set up using the test bench source and this source must be very accurate. The ADE7754 should be configured for Active Energy measurement mode. An Active Energy measurement is first made with PF=1 (measurement A). A second Active Energy measurement should be made at PF=0.5 (measurement B). The frequency output CF can be used for this measurement. Using the formula shown below the phase error is easily calculated:

$$\text{Phase Error } (^{\circ}) = \tan^{-1} \left(\frac{B - A/2}{A/2 \cdot \sqrt{3}} \right)$$

For example, using the frequency output CF to measure power, a frequency of 3.66621Hz is recorded for a PF=1. The PF is then set to 0.5 lag and a measurement of 1.83817Hz is obtained. Using the formula above the phase error on Channel 1 is calculated as:

$$\text{Phase Error } (^{\circ}) = \tan^{-1} \left(\frac{1.83817 - 3.66621/2}{3.66621/2 \cdot \sqrt{3}} \right) = +0.091^{\circ}$$

The formula will also give the correct sign for the phase error. In this example the phase error is calculated as +0.091° at the input to the Current channel of ADE7754. This means that the CT has introduced a phase lead of 0.091°. Therefore the phase difference at the input to the Current channel is now 59.89° lag instead of 60° lag. Determining whether the error is a lead or lag can also be figured intuitively from the frequency output. Figure 14 shows how the output frequency varies with phase ($\cos\{\phi\}$). Since the output frequency B (1.83817Hz) at the PF=0.5 lag setting in the example is actually greater than A/2 (1.833105Hz), this means the phase error between the current channel and the voltage channel was actually less than 60°. This means there was additional lead in the current channel due to the CT.




Figure 14—CF Frequency Vs Phase(PF)

Using the Phase Calibration to correct small (<0.5°) external phase errors

From the previous example it is seen that the CT introduced a phase lead in the current channel of 0.091°. Therefore instead of 60° phase difference between the current channel and the voltage channel, it is actually 59.89°. In order to bring the phase difference back to 60°, the phase compensation circuit in the voltage channel is used to introduce an extra lead of 0.091°. This is achieved by reducing the amount of time delay in the voltage channel.

The maximum time delay adjustment in the voltage channel is $\pm 38.4\mu\text{s}$ with a CLKIN of 10MHz. The PHCAL registers are signed 2's complement 5 bit registers. Therefore each LSB is equivalent to 1.2μs. In this example the line frequency is 50Hz. This means each LSB is equivalent to $(360^{\circ} \times 1.2\mu\text{s} \times 50) = 0.022^{\circ}$. To introduce a lead of 0.091° the delay in the voltage channel must be reduced. This is achieved by writing -4 (1Ch) or +0.088° to the PHCAL register.




Figure 15—Writing to the PHCAL register to correct phase error

Correcting large external phase errors

In this example the phase correction range at 50Hz is only approximately $\pm 0.7^{\circ}$. However it is best to only use the PHCAL register for small phase corrections, i.e., $< 0.5^{\circ}$. If larger corrections are required the larger part of the correction can be made using external passive component. For example the resistors in the anti-alias filter can be modified to shift the corner frequency of the filter so as to introduce more or less lag. The lag through the anti-alias filters with 1kΩ and 33nF is 0.56° at 50Hz. Fine adjustments can be made with the PHCAL register. Note that typically CT phase shift will not vary significantly from part to part. If a CT phase shift is 1°, then the part to part variation should only be about $\pm 0.1^{\circ}$. Therefore the bulk of the phase shift (1°) can be canceled with fixed component values at design. The remaining small adjustments can be made in production using the PHCAL registers.


EVAL-ADE7754EB

EvaluationboardBOM

Designator	Value	Description
R4, R24, R26, R29-R31 R5, R7, R9, R13, R15, R19, R21, R32-R36, R50-51	10k Ω , 5%, 1/4W	Resistor, no special requirements 100 Ω , 5%, 1/4W Resistor, no special requirements
R6, R11-12, R17-18, R23, R52	51 Ω , 5%, 1/4W	Not placed unless external clock is being used
R8, R10, R14, R16, R20, R22, R40, R43-44, R47	1k Ω , 0.1%, 1/4W	± 15 ppm/ $^{\circ}$ C Resistor, good tolerance, used as part of the analog filter network. These resistors are not soldered, but are plugged into PCB pin sockets for easy modification by the customer. Low drift WELWYN RC55 Series, FARNELL part no. 339-179
R25	10 Ω , 5%, 1/4W	Resistor, no special requirements
R38-39	20 Ω , 5%, 1/4W	Resistor, no special requirements
R41-42, R45-46, R48-49	499k Ω , 0.1%, 1/4W	Pin socketed, ± 15 ppm/ $^{\circ}$ C Low drift, WELWYN RC55 Series. Farnell part no. 338-484
R53-56, R61, R63-68	820 Ω , 5%, 1/4W	Resistor, no special requirements
R69	0 Ω , 10%, 1/4W	
C1-16	33nF, 10%, 50 volt	X7R Capacitor, part of the filter network. These resistors are not soldered, but are plugged into PCB mount sockets for easy modification by the customer. SR15 series AVX-KYOCERNA, FARNELL part no. 108-948C5,
C17	220pF	PANASONIC ECQ-P1H221JZ
C18-23	10uF, Tantalum	Power supply decoupling capacitors, 10%, 16V
C25, C39	22pF, ceramic	Gate oscillator load capacitors, FARNELL part no. 108-927
C24, C26-34, C37-38	100nF, 25V	Power supply decoupling capacitors, 10%, X7R type, AVX-KYOCERNA, FARNELL part no. 108-950
C11, C21, C50, C51, CR1-3	LED	Low current, Red, FARNELL part no. 637-087
CR4	Diode	1N4001, Motorola
J1-3	BNC connector	Straight square, 1.3mm holes, 10.2mm x 10.2mm FARNELL part no. 149-453
JP1-29, JP32, JP35, JP38, JP46-49	2 Pin header	2-Pin, 0.025 Sq., 0.01 Ctrs, Compnt Corp., CSS-02-02
JP43A-B, JP44A-B, P9	2 Pin header x 2 3 Pin header x 2	2-Pin, 0.025 Sq., 0.01 Ctrs, Compnt Corp., CSS-02-02U1 2-Pin, 0.025 Sq., 0.01 Ctrs, Compnt Corp., CSS-02-02U1
	ADE7754	Supplied by Analog Devices Inc.
U1-2	74HC08	Quad CMOS AND gates
U3-6	HCPL2232	HP Optical Isolator
U7	AD780	2.5V reference, Supplied by Analog Devices Inc.
XTAL	10 MHz	Quartz Crystal, HC-49(US)
P1	D-Sub 25 way male	AMP 747238-4 Right angle "D-Sub" 8mm PCB mount, DigiKey no. 747238-4


PRELIMINARY TECHNICAL DATA**EVAL-ADE7754EB**

P2-8	screw terminal	15A, 2.5mm cable screw terminal sockets. WEILAND part no. 25.161.0253
P10	screw terminal	15A, 2.5mm cable screw terminal sockets. WEILAND part no. 25.161.0353



EVAL-ADE7754EB

PCB layout - Component Side



PCB layout - Solder Side

